
Midmark M11 - v1.0.5
Total Cycles: 2246

Name: RITTER2
Sterilizer ID: V2313228
04/13/2023 11:24 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.8 F 0.0
2:00 162.6 F 3.0
4:00 172.2 F 6.7
6:00 198.3 F 11.6
8:00 216.0 F 5.6
10:00 227.8 F 10.7
12:00 230.4 F 7.5
14:00 242.7 F 13.3
16:00 242.2 F 12.1
18:00 254.8 F 18.5
20:00 265.8 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.3
1:00 272.6 F 30.1
2:00 272.1 F 29.5
3:00 271.2 F 29.4
4:00 272.8 F 30.2

Min 271.2 F 29.1
Max 272.8 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.5 F 0.0
5:00 210.9 F 0.0
10:00 223.5 F 0.0
15:00 217.7 F 0.0
20:00 221.6 F 0.0
25:00 231.5 F 0.0
30:00 222.4 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 21:47
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:11
CYCLE COMPLETE

04/13/2023 12:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____