
Midmark M11 - v1.0.0
Total Cycles: 1265

Name: RITTER 1
Sterilizer ID: V2178198
10/14/2021 09:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.6 F 0.0
2:00 195.4 F 3.1
4:00 220.3 F 9.6
6:00 226.5 F 6.1
8:00 238.5 F 12.0
10:00 238.8 F 10.2
12:00 237.9 F 9.1
14:00 253.1 F 16.7
16:00 265.7 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 271.9 F 28.3
2:00 272.6 F 28.7
3:00 272.2 F 28.8
4:00 271.8 F 28.2

Min 271.5 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.6 F 0.0
5:00 215.1 F 0.0
10:00 223.4 F 0.0
15:00 227.4 F 0.0
20:00 210.1 F 0.0
25:00 202.4 F 0.0
30:00 199.8 F 0.0
Drying Complete

FILLING: 2:34
HEATING: 17:31
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:57:11
CYCLE COMPLETE

10/14/2021 10:25 AM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____