

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 006601  
14:07 13/11/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:06  
PRESSURIZING 2:08  
STERILIZING 3:11  
136.4 C 223kPa 3:11  
136.4 C 224kPa 3:41  
136.4 C 224kPa 4:11  
136.4 C 222kPa 4:41  
136.4 C 223kPa 5:11  
136.3 C 224kPa 5:41  
136.3 C 222kPa 6:11  
136.3 C 222kPa 6:41  
136.3 C 222kPa 6:41  
VENTING 6:41  
AIR DRYING 7:30  
CYCLE COMPLETE 7:54

-----  
Digital Signature #  
42EA270422BF1C4D  
-----