

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1399

Name: RITTER 1  
Sterilizer ID: V2178198  
11/29/2021 10:48 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 144.8 F 0.0  
2:00 150.1 F 1.1  
4:00 171.8 F 4.4  
6:00 192.0 F 8.5  
8:00 210.3 F 13.6  
10:00 221.7 F 7.3  
12:00 234.3 F 12.3  
14:00 235.8 F 9.3  
16:00 237.6 F 9.3  
18:00 247.4 F 14.0  
20:00 259.8 F 20.7  
22:00 270.9 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.7  
1:00 272.8 F 29.2  
2:00 271.4 F 28.7  
3:00 272.7 F 29.0  
4:00 271.9 F 29.1

Min 271.3 F 28.5  
Max 272.9 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.2 F 0.0  
5:00 219.2 F 0.0  
10:00 216.7 F 0.0  
15:00 208.1 F 0.0  
20:00 221.5 F 0.0  
25:00 231.8 F 0.0  
30:00 217.5 F 0.0  
Drying Complete

FILLING: 1:17  
HEATING: 23:25  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:01:57  
CYCLE COMPLETE

11/29/2021 11:53 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02