

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3814

Name: RITTER2  
Sterilizer ID: V2313228  
07/25/2024 12:00 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 104.3 F 0.0  
2:00 105.5 F 0.0  
4:00 107.1 F 0.1  
6:00 139.5 F 1.8  
8:00 179.6 F 4.8  
10:00 198.9 F 8.4  
12:00 215.0 F 12.1  
14:00 222.9 F 6.2  
16:00 228.8 F 10.4  
18:00 235.0 F 8.4  
20:00 239.6 F 12.0  
22:00 238.1 F 10.2  
24:00 249.8 F 16.1  
26:00 260.8 F 22.1  
28:00 270.6 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.8  
1:00 272.5 F 30.0  
2:00 272.6 F 30.4  
3:00 271.7 F 29.8  
4:00 272.6 F 30.2

Min 271.7 F 29.7  
Max 272.7 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.4 F 0.0  
5:00 217.2 F 0.0  
10:00 226.5 F 0.0  
15:00 230.3 F 0.0  
20:00 230.7 F 0.0  
25:00 224.8 F 0.0  
30:00 222.8 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 29:33  
STERILIZING: 4:00  
VENTING: 3:40  
DRYING: 30:00  
TOTAL CYCLE: 01:08:04  
CYCLE COMPLETE

07/25/2024 01:08 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_