
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.9uS / 0.5ppm
CYCLE NUMBER 001470
11:01 10/04/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:03
PRESSURIZING 6:50
STERILIZING 8:29
136.1 C 220kPa 8:29
136.4 C 222kPa 8:59
136.3 C 221kPa 9:29
136.2 C 221kPa 9:59
136.4 C 222kPa 10:29
136.4 C 219kPa 10:59
136.4 C 220kPa 11:29
136.4 C 222kPa 11:59
VENTING 11:59
AIR DRYING 13:05
CYCLE COMPLETE 28:23

Digital Signature #
587959B88E17D67F
