
Midmark M11 - v1.0.0
Total Cycles: 3160

Name: RITTER 1
Sterilizer ID: V2178198
04/25/2023 10:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.3 F 0.0
2:00 194.7 F 0.0
4:00 214.4 F 3.1
6:00 229.5 F 8.2
8:00 241.0 F 13.5
10:00 240.1 F 11.2
12:00 238.6 F 9.7
14:00 238.1 F 8.5
16:00 252.5 F 16.3
18:00 265.1 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.3 F 28.4
2:00 272.1 F 28.6
3:00 272.1 F 28.2
4:00 272.1 F 28.6

Min 271.5 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.7 F 0.0
5:00 222.8 F 0.0
10:00 227.1 F 0.0
15:00 225.2 F 0.0
20:00 228.1 F 0.0
25:00 229.5 F 0.0
30:00 229.2 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 19:42
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:58:10
CYCLE COMPLETE

04/25/2023 11:40 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____