
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000128
11:08 08/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:36
PRESSURIZING 4:36
STERILIZING 5:55
136.0 C 221kPa 5:55
136.5 C 226kPa 6:25
136.3 C 224kPa 6:55
136.4 C 223kPa 7:25
136.4 C 226kPa 7:55
136.3 C 223kPa 8:25
136.4 C 222kPa 8:55
136.2 C 224kPa 9:25
136.2 C 224kPa 9:25
VENTING 9:25
AIR DRYING 10:28
CYCLE COMPLETE 10:33

Digital Signature #
9EABEE0100C5C02D
