
Midmark M11 - v1.0.0
Total Cycles: 1936

Name: RITTER 1
Sterilizer ID: V2178198
05/12/2022 07:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.9 F 0.0
2:00 94.7 F 0.0
4:00 91.9 F 0.0
6:00 124.6 F 0.3
8:00 154.5 F 1.9
10:00 179.7 F 4.0
12:00 201.9 F 7.1
14:00 217.8 F 10.8
16:00 221.9 F 5.7
18:00 228.6 F 7.5
20:00 239.0 F 11.7
22:00 234.1 F 8.0
24:00 245.6 F 12.4
26:00 244.0 F 11.9
28:00 255.8 F 18.1
30:00 266.4 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 272.4 F 28.4
2:00 272.1 F 28.5
3:00 272.1 F 28.1
4:00 272.4 F 28.8

Min 271.6 F 27.8
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.5 F 0.0
5:00 203.9 F 0.0
10:00 215.8 F 0.0
15:00 233.4 F 0.0
20:00 203.9 F 0.0
25:00 202.9 F 0.0
30:00 184.2 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 31:36
STERILIZING: 4:00
VENTING: 3:00
DRYING: 30:00
TOTAL CYCLE: 01:09:24
CYCLE COMPLETE

05/12/2022 08:51 AM
ACCEPTED: X
REJECTED:
Operator: U02