

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3113

Name: RITTER 1  
Sterilizer ID: V2178198  
04/14/2023 10:06 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 161.9 F 0.0  
2:00 163.6 F 0.0  
4:00 176.0 F 3.1  
6:00 202.9 F 7.4  
8:00 222.6 F 12.5  
10:00 226.1 F 7.2  
12:00 238.4 F 12.0  
14:00 235.6 F 9.0  
16:00 240.2 F 9.9  
18:00 246.3 F 13.5  
20:00 258.3 F 20.0  
22:00 269.0 F 26.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.5  
1:00 272.6 F 29.2  
2:00 271.7 F 28.4  
3:00 272.1 F 29.2  
4:00 272.3 F 29.0

Min 271.4 F 28.3  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.1 F 0.0  
5:00 215.1 F 0.0  
10:00 225.7 F 0.0  
15:00 233.3 F 0.0  
20:00 231.0 F 0.0  
25:00 225.8 F 0.0  
30:00 225.6 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 23:08  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:01:12  
CYCLE COMPLETE

04/14/2023 11:07 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_