
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002386
11:19 16/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:55
PRESSURIZING 2:00
STERILIZING 3:04
136.4 C 225kPa 3:04
136.2 C 224kPa 3:34
136.2 C 225kPa 4:04
136.4 C 224kPa 4:34
136.2 C 223kPa 5:04
136.1 C 223kPa 5:34
136.4 C 226kPa 6:04
136.3 C 220kPa 6:34
136.3 C 220kPa 6:34
VENTING 6:34
AIR DRYING 7:25
CYCLE COMPLETE 7:52

Digital Signature #
5D83F28DC9E660EC
