

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001468  
7:42 10/01/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 5:15  
PRESSURIZING 7:09  
STERILIZING 8:54  
136.0 C 220kPa 8:54  
136.4 C 222kPa 9:24  
136.4 C 222kPa 9:54  
136.4 C 222kPa 10:24  
136.4 C 221kPa 10:54  
136.2 C 221kPa 11:24  
136.4 C 219kPa 11:54  
136.3 C 220kPa 12:24  
VENTING 12:24  
AIR DRYING 13:29  
CYCLE COMPLETE 43:29

-----  
Digital Signature #  
C6A4CBACB262FC9B  
-----