
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001351
16:35 16/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:49
PRESSURIZING 2:56
STERILIZING 4:03
136.0 C 220kPa 4:03
136.2 C 224kPa 4:33
136.4 C 225kPa 5:03
136.3 C 225kPa 5:33
136.4 C 224kPa 6:03
136.1 C 222kPa 6:33
136.4 C 225kPa 7:03
136.4 C 222kPa 7:33
136.4 C 222kPa 7:33
VENTING 7:33
AIR DRYING 8:24
CYCLE COMPLETE 13:12

Digital Signature #
B9FBA5DA39C4CDF0
