
Midmark M11 - v1.0.5
Total Cycles: 728

Name: RITTER2
Sterilizer ID: V2313228
01/26/2022 07:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.2 F 0.0
2:00 77.8 F 0.0
4:00 107.8 F 0.0
6:00 150.3 F 0.1
8:00 173.0 F 1.7
10:00 186.4 F 4.6
12:00 186.5 F 8.3
14:00 201.9 F 12.1
16:00 216.5 F 5.1
18:00 226.7 F 9.1
20:00 237.3 F 13.5
22:00 237.8 F 10.5
24:00 236.6 F 8.5
26:00 248.5 F 14.7
28:00 259.8 F 21.0
30:00 270.1 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.6 F 29.6
2:00 272.3 F 29.0
3:00 271.5 F 28.7
4:00 272.6 F 29.6

Min 271.3 F 28.5
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.6 F 0.0
5:00 204.5 F 0.0
10:00 210.2 F 0.0
15:00 210.7 F 0.0
20:00 209.5 F 0.0
25:00 208.8 F 0.0
30:00 211.1 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 30:59
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:09:00
CYCLE COMPLETE

01/26/2022 08:48 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____