
Midmark M11 - v1.0.0
Total Cycles: 4744

Name: RITTER 1
Sterilizer ID: V2178198
07/18/2024 10:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 128.9 F 0.0
2:00 141.0 F 1.3
4:00 143.2 F 4.1
6:00 167.7 F 7.8
8:00 201.9 F 12.1
10:00 217.9 F 4.2
12:00 228.9 F 8.5
14:00 239.3 F 12.9
16:00 235.8 F 9.2
18:00 239.9 F 10.1
20:00 244.3 F 12.4
22:00 256.0 F 18.5
24:00 266.6 F 25.0
26:00 272.9 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.8
1:00 272.9 F 29.2
2:00 272.4 F 29.3
3:00 271.4 F 28.7
4:00 272.6 F 29.0

Min 271.3 F 28.6
Max 273.1 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.5 F 0.0
5:00 233.6 F 0.0
10:00 225.5 F 0.0
15:00 228.6 F 0.0
20:00 233.4 F 0.0
25:00 234.5 F 0.0
30:00 226.9 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 26:08
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:04:23
CYCLE COMPLETE

07/18/2024 11:12 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____