

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2497

Name: RITTER2  
Sterilizer ID: V2313228  
06/12/2023 08:22 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.2 F 0.0  
2:00 73.0 F 0.0  
4:00 81.7 F 0.0  
6:00 105.4 F 0.0  
8:00 145.0 F 0.1  
10:00 176.4 F 3.1  
12:00 189.0 F 6.5  
14:00 207.8 F 10.1  
16:00 224.6 F 9.6  
18:00 225.4 F 8.1  
20:00 237.1 F 12.3  
22:00 235.5 F 9.3  
24:00 239.5 F 10.3  
26:00 245.7 F 13.5  
28:00 256.9 F 19.5  
30:00 267.0 F 25.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.9  
1:00 272.6 F 29.7  
2:00 271.6 F 28.8  
3:00 272.2 F 29.6  
4:00 272.3 F 29.3

Min 271.2 F 28.7  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.9 F 0.0  
5:00 213.9 F 0.0  
10:00 209.8 F 0.0  
15:00 198.3 F 0.0  
20:00 216.0 F 0.0  
25:00 228.1 F 0.0  
30:00 213.7 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 31:35  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:09:34  
CYCLE COMPLETE

06/12/2023 09:31 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_