
Midmark M11 - v1.0.5
Total Cycles: 2616

Name: RITTER2
Sterilizer ID: V2313228
07/06/2023 10:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.3 F 0.0
2:00 169.6 F 2.5
4:00 187.9 F 6.7
6:00 206.6 F 12.0
8:00 222.8 F 7.5
10:00 235.5 F 13.0
12:00 237.7 F 10.7
14:00 235.9 F 9.2
16:00 250.2 F 16.0
18:00 261.7 F 22.7
20:00 272.4 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.7 F 30.0
2:00 271.8 F 29.2
3:00 271.5 F 29.5
4:00 272.8 F 30.0

Min 271.2 F 29.0
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.3 F 0.0
5:00 217.1 F 0.0
10:00 227.1 F 0.0
15:00 231.5 F 0.0
20:00 228.4 F 0.0
25:00 230.2 F 0.0
30:00 227.3 F 0.0
Drying Complete

FILLING: 1:43
HEATING: 20:29
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:26
CYCLE COMPLETE

07/06/2023 11:08 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____