
Midmark M11 - v1.0.0
Total Cycles: 5040

Name: RITTER 1
Sterilizer ID: V2178198
10/31/2024 08:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.9 F 0.0
2:00 80.7 F 0.0
4:00 90.4 F 0.0
6:00 129.7 F 0.2
8:00 160.7 F 0.3
10:00 184.9 F 0.8
12:00 200.9 F 4.1
14:00 211.6 F 7.4
16:00 221.9 F 10.8
18:00 224.5 F 6.4
20:00 234.8 F 9.6
22:00 238.7 F 9.9
24:00 243.9 F 12.3
26:00 241.3 F 10.7
28:00 254.3 F 17.3
30:00 265.7 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.9
1:00 272.7 F 28.6
2:00 271.5 F 27.8
3:00 272.7 F 28.8
4:00 272.1 F 28.1

Min 271.5 F 27.8
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 173.0 F 0.0
5:00 190.5 F 0.0
10:00 195.8 F 0.0
15:00 192.8 F 0.0
20:00 192.3 F 0.0
25:00 195.4 F 0.0
30:00 192.8 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 31:46
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:09:53
CYCLE COMPLETE

10/31/2024 09:53 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____