

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001236  
12:34 31/08/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:02  
PRESSURIZING 4:04  
STERILIZING 5:28  
136.2 C 220kPa 5:28  
136.2 C 222kPa 5:58  
136.2 C 223kPa 6:28  
136.2 C 222kPa 6:58  
136.4 C 225kPa 7:28  
136.1 C 222kPa 7:58  
136.4 C 224kPa 8:28  
136.4 C 223kPa 8:58  
136.4 C 223kPa 8:58  
VENTING 8:58  
AIR DRYING 9:54  
CYCLE COMPLETE 24:54  
-----

Digital Signature #  
CF8A28464D970BF7  
-----