

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2717

Name: RITTER 1  
Sterilizer ID: V2178198  
12/22/2022 10:46 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 196.1 F 0.0  
2:00 191.3 F 2.1  
4:00 192.1 F 6.9  
6:00 212.6 F 12.7  
8:00 227.5 F 7.5  
10:00 242.5 F 13.2  
12:00 241.2 F 10.9  
14:00 238.4 F 9.4  
16:00 253.2 F 16.6  
18:00 265.2 F 23.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.0  
1:00 272.1 F 28.4  
2:00 271.9 F 28.5  
3:00 271.8 F 28.2  
4:00 272.0 F 28.6  
  
Min 271.2 F 27.8  
Max 272.6 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.3 F 0.0  
5:00 224.4 F 0.0  
10:00 241.9 F 0.0  
15:00 239.8 F 0.0  
20:00 228.7 F 0.0  
25:00 228.2 F 0.0  
30:00 227.9 F 0.0  
Drying Complete

FILLING: 1:54  
HEATING: 19:44  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:58:50  
CYCLE COMPLETE

12/22/2022 11:44 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_