
Midmark M11 - v1.0.5
Total Cycles: 669

Name: RITTER2
Sterilizer ID: V2313228
12/30/2021 09:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 129.0 F 0.0
2:00 126.6 F 1.1
4:00 138.1 F 3.7
6:00 169.1 F 7.7
8:00 194.6 F 12.6
10:00 217.8 F 6.0
12:00 236.0 F 10.9
14:00 231.7 F 7.7
16:00 244.1 F 12.7
18:00 243.4 F 12.4
20:00 256.1 F 18.9
22:00 267.2 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.8
1:00 273.0 F 30.0
2:00 271.8 F 29.1
3:00 272.7 F 30.0
4:00 272.5 F 29.4

Min 271.0 F 28.6
Max 273.2 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.1 F 0.0
5:00 236.9 F 0.0
10:00 236.3 F 0.0
15:00 211.6 F 0.0
20:00 211.5 F 0.0
25:00 213.9 F 0.0
30:00 205.9 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 23:27
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 01:01:34
CYCLE COMPLETE

12/30/2021 10:21 AM
ACCEPTED: X
REJECTED:
Operator: U02