
Midmark M11 - v1.0.5
Total Cycles: 1288

Name: RITTER2
Sterilizer ID: V2313228
07/13/2022 09:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 209.0 F 0.0
2:00 198.8 F 1.5
4:00 215.6 F 6.6
6:00 225.3 F 12.0
8:00 232.2 F 8.6
10:00 240.7 F 10.9
12:00 244.1 F 13.0
14:00 243.7 F 12.3
16:00 256.9 F 19.3
18:00 268.3 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 271.9 F 29.1
2:00 272.1 F 29.6
3:00 271.6 F 28.7
4:00 272.3 F 29.5

Min 271.4 F 28.6
Max 272.4 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.2 F 0.0
5:00 222.4 F 0.0
10:00 229.3 F 0.0
15:00 224.5 F 0.0
20:00 223.9 F 0.0
25:00 222.3 F 0.0
30:00 220.1 F 0.0

Drying Complete

FILLING: 1:46
HEATING: 19:12
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:58:02
CYCLE COMPLETE

07/13/2022 10:24 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____