

-----  
Surrey, BC  
Canada  
-----

STATIM 5000 S501R711  
SN 511018F00076  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000049  
8:43 14/09/2020  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 6:18  
PRESSURIZING 8:42  
STERILIZING 10:50  
136.0 C 223kPa 10:50  
136.5 C 227kPa 11:20  
136.5 C 227kPa 11:50  
136.4 C 226kPa 12:20  
136.1 C 225kPa 12:50  
136.5 C 224kPa 13:20  
136.2 C 226kPa 13:51  
136.4 C 227kPa 14:20  
VENTING 14:20  
AIR DRYING 15:33  
CYCLE COMPLETE 45:33

-----  
Digital Signature #  
53D2A162543EB8E7  
-----