

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5318

Name: RITTER 1  
Sterilizer ID: V2178198  
02/05/2025 03:42 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 155.4 F 0.0  
2:00 158.1 F 3.1  
4:00 168.9 F 6.8  
6:00 191.8 F 11.1  
8:00 209.9 F 4.3  
10:00 222.1 F 9.1  
12:00 231.9 F 10.4  
14:00 237.1 F 10.3  
16:00 235.8 F 8.0  
18:00 247.8 F 14.3  
20:00 259.5 F 20.7  
22:00 270.2 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.5  
1:00 272.8 F 29.3  
2:00 271.6 F 28.4  
3:00 272.8 F 29.2  
4:00 271.7 F 28.4

Min 271.4 F 28.4  
Max 272.9 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.7 F 0.0  
5:00 199.2 F 0.0  
10:00 222.9 F 0.0  
15:00 221.8 F 0.0  
20:00 225.9 F 0.0  
25:00 220.4 F 0.0  
30:00 230.5 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 22:45  
STERILIZING: 4:00  
VENTING: 3:38  
DRYING: 30:00  
TOTAL CYCLE: 01:01:19  
CYCLE COMPLETE

02/05/2025 04:43 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_