
Midmark M11 - v1.0.0
Total Cycles: 2034

Name: RITTER 1
Sterilizer ID: V2178198
06/20/2022 09:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 128.6 F 0.0
2:00 127.9 F 0.1
4:00 135.0 F 1.4
6:00 171.4 F 3.7
8:00 204.6 F 7.3
10:00 213.7 F 12.1
12:00 223.5 F 5.9
14:00 232.2 F 10.5
16:00 234.1 F 7.7
18:00 243.4 F 12.6
20:00 241.0 F 10.8
22:00 253.8 F 17.2
24:00 264.9 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.6 F 28.8
2:00 271.3 F 28.2
3:00 272.6 F 29.0
4:00 271.7 F 28.2

Min 271.3 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.3 F 0.0
5:00 221.7 F 0.0
10:00 219.4 F 0.0
15:00 222.4 F 0.0
20:00 233.7 F 0.0
25:00 235.0 F 0.0
30:00 232.8 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 25:57
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:03:51
CYCLE COMPLETE

06/20/2022 10:14 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____