
Midmark M11 - v1.0.0
Total Cycles: 837

Name: RITTER 1
Sterilizer ID: V2178198
05/26/2021 11:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.0 F 0.0
2:00 194.4 F 3.3
4:00 198.8 F 7.6
6:00 214.1 F 12.1
8:00 222.3 F 6.6
10:00 234.2 F 11.4
12:00 234.3 F 8.6
14:00 241.9 F 11.2
16:00 246.1 F 13.4
18:00 257.9 F 19.8
20:00 268.7 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.6 F 29.2
2:00 271.9 F 28.6
3:00 271.9 F 28.9
4:00 272.6 F 29.2

Min 271.5 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.4 F 0.0
5:00 231.0 F 0.0
10:00 217.5 F 0.0
15:00 209.8 F 0.0
20:00 226.6 F 0.0
25:00 218.9 F 0.0
30:00 222.2 F 0.0
Drying Complete

FILLING: 1:38
HEATING: 21:14
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:00:25
CYCLE COMPLETE

05/26/2021 12:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____