

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2016

Name: RITTER2  
Sterilizer ID: V2313228  
02/07/2023 08:21 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 77.8 F 0.0  
2:00 79.6 F 0.0  
4:00 99.1 F 0.0  
6:00 114.7 F 0.0  
8:00 183.5 F 0.0  
10:00 192.2 F 2.0  
12:00 208.8 F 5.4  
14:00 224.8 F 9.0  
16:00 232.9 F 13.1  
18:00 232.0 F 8.9  
20:00 241.3 F 11.4  
22:00 242.4 F 12.3  
24:00 241.1 F 11.2  
26:00 254.1 F 17.9  
28:00 265.8 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.8  
1:00 272.6 F 29.6  
2:00 271.7 F 28.8  
3:00 271.8 F 29.3  
4:00 272.5 F 29.5  
  
Min 271.3 F 28.6  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.6 F 0.0  
5:00 201.9 F 0.0  
10:00 217.9 F 0.0  
15:00 229.1 F 0.0  
20:00 232.7 F 0.0  
25:00 226.6 F 0.0  
30:00 227.5 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 29:43  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:07:38  
CYCLE COMPLETE

02/07/2023 09:28 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_