
Midmark M11 - v1.0.0
Total Cycles: 2126

Name: RITTER 1
Sterilizer ID: V2178198
07/14/2022 08:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.1 F 0.0
2:00 134.2 F 0.0
4:00 142.4 F 1.3
6:00 173.1 F 4.9
8:00 196.0 F 9.1
10:00 207.1 F 9.4
12:00 224.1 F 7.5
14:00 236.4 F 12.0
16:00 234.3 F 8.6
18:00 242.6 F 11.3
20:00 243.6 F 12.3
22:00 255.6 F 18.5
24:00 266.5 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.5 F 29.0
2:00 271.5 F 28.7
3:00 272.2 F 28.9
4:00 271.5 F 28.7

Min 271.2 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.9 F 0.0
5:00 227.5 F 0.0
10:00 218.4 F 0.0
15:00 202.2 F 0.0
20:00 208.5 F 0.0
25:00 210.6 F 0.0
30:00 210.6 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 25:35
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:03:36
CYCLE COMPLETE

07/14/2022 09:37 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____