
Midmark M11 - v1.0.5
Total Cycles: 2017

Name: RITTER2
Sterilizer ID: V2313228
02/07/2023 09:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.5 F 0.0
2:00 158.9 F 0.0
4:00 173.1 F 0.7
6:00 192.4 F 5.0
8:00 214.7 F 9.8
10:00 229.6 F 7.4
12:00 237.1 F 10.8
14:00 235.5 F 8.6
16:00 240.2 F 10.2
18:00 248.8 F 14.8
20:00 261.5 F 22.0
22:00 272.8 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.5 F 29.3
2:00 271.4 F 28.6
3:00 272.5 F 29.5
4:00 271.8 F 28.9

Min 271.3 F 28.5
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.4 F 0.0
5:00 228.8 F 0.0
10:00 236.4 F 0.0
15:00 227.8 F 0.0
20:00 228.0 F 0.0
25:00 229.2 F 0.0
30:00 228.7 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 22:23
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:00:17
CYCLE COMPLETE

02/07/2023 10:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____