

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1224

Name: RITTER 1  
Sterilizer ID: V2178198  
09/30/2021 12:14 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 134.1 F 0.0  
2:00 151.1 F 2.3  
4:00 178.1 F 5.8  
6:00 200.5 F 11.0  
8:00 216.0 F 5.2  
10:00 230.8 F 10.4  
12:00 231.1 F 7.1  
14:00 244.3 F 13.2  
16:00 242.7 F 11.7  
18:00 256.2 F 18.6  
20:00 267.7 F 25.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.3  
1:00 272.7 F 29.0  
2:00 271.6 F 28.2  
3:00 272.7 F 29.2  
4:00 272.3 F 28.7  
  
Min 271.5 F 28.2  
Max 272.8 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.7 F 0.0  
5:00 219.5 F 0.0  
10:00 229.7 F 0.0  
15:00 228.8 F 0.0  
20:00 220.0 F 0.0  
25:00 221.6 F 0.0  
30:00 218.5 F 0.0  
Drying Complete

FILLING: 1:39  
HEATING: 21:19  
STERILIZING: 4:00  
VENTING: 3:00  
DRYING: 30:00  
TOTAL CYCLE: 00:59:58  
CYCLE COMPLETE

09/30/2021 01:13 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_