
Midmark M11 - v1.0.0
Total Cycles: 4682

Name: RITTER 1
Sterilizer ID: V2178198
06/26/2024 04:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.7 F 0.0
2:00 180.8 F 4.5
4:00 193.4 F 9.3
6:00 216.0 F 6.9
8:00 227.6 F 9.4
10:00 232.9 F 8.8
12:00 242.3 F 12.7
14:00 241.7 F 11.6
16:00 255.3 F 18.5
18:00 267.2 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.3 F 29.1
2:00 271.7 F 29.1
3:00 272.1 F 29.0
4:00 271.6 F 29.0

Min 271.3 F 28.5
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.5 F 0.0
5:00 221.0 F 0.0
10:00 230.0 F 0.0
15:00 226.0 F 0.0
20:00 219.9 F 0.0
25:00 231.1 F 0.0
30:00 228.9 F 0.0

Drying Complete

FILLING: 1:36
HEATING: 19:23
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 00:58:31
CYCLE COMPLETE

06/26/2024 05:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____