

-----  
Midmark M11 - v1.0.0  
Total Cycles: 583

Name: MIDMARK  
Sterilizer ID: V2178198  
12/29/2020 07:58 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	71.5 F	0.0
2:00	72.2 F	0.0
4:00	83.5 F	0.0
6:00	108.7 F	0.1
8:00	145.7 F	0.2
10:00	177.7 F	0.3
12:00	193.4 F	1.0
14:00	208.6 F	3.8
16:00	216.1 F	6.3
18:00	225.5 F	9.3
20:00	232.9 F	12.9
22:00	233.3 F	8.5
24:00	243.5 F	13.2
26:00	241.7 F	11.0
28:00	238.6 F	9.3
30:00	251.7 F	15.6
32:00	262.4 F	21.8
34:00	272.2 F	28.2

STERILIZING

mm:ss	Degrees	PSI
0:00	271.5 F	27.8
1:00	272.3 F	28.2
2:00	271.8 F	28.0
3:00	272.1 F	28.1
4:00	271.9 F	28.1

Min 271.4 F 27.6  
Max 272.6 F 28.5

VENTING CHAMBER

DRYING

Drying Start	mm:ss	Degrees	PSI
	0:00	213.2 F	0.0
	5:00	200.2 F	0.0
	10:00	195.2 F	0.0
	15:00	209.2 F	0.0
	20:00	212.5 F	0.0
Drying Complete	25:00	215.7 F	0.0
	30:00	214.7 F	0.0

FILLING: 0:51  
HEATING: 34:28  
STERILIZING: 4:00  
VENTING: 3:25  
DRYING: 30:00  
TOTAL CYCLE: 01:12:45  
CYCLE COMPLETE

12/29/2020 09:10 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_