
Midmark M11 - v1.0.0
Total Cycles: 4193

Name: RITTER 1
Sterilizer ID: V2178198
01/11/2024 10:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.4 F 0.0
2:00 151.9 F 1.5
4:00 156.7 F 4.3
6:00 182.7 F 8.0
8:00 207.4 F 12.3
10:00 221.4 F 5.6
12:00 234.0 F 9.8
14:00 237.5 F 10.2
16:00 239.8 F 10.9
18:00 235.0 F 8.3
20:00 248.4 F 14.4
22:00 259.1 F 20.4
24:00 269.3 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.6 F 29.0
2:00 272.4 F 29.4
3:00 271.5 F 28.7
4:00 272.6 F 29.3

Min 271.3 F 28.7
Max 273.0 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.5 F 0.0
5:00 231.7 F 0.0
10:00 210.9 F 0.0
15:00 224.8 F 0.0
20:00 219.4 F 0.0
25:00 216.1 F 0.0
30:00 202.9 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 25:49
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:04:42
CYCLE COMPLETE

01/11/2024 11:34 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____