
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000120
16:24 03/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:23
PRESSURIZING 4:24
STERILIZING 5:32
136.5 C 225kPa 5:32
136.4 C 226kPa 6:02
136.3 C 225kPa 6:32
136.3 C 225kPa 7:02
136.3 C 225kPa 7:32
136.3 C 225kPa 8:02
136.4 C 224kPa 8:32
136.4 C 225kPa 9:02
136.4 C 225kPa 9:02
VENTING 9:02
AIR DRYING 10:03
CYCLE COMPLETE 23:42

Digital Signature #
96C90B596B7018A7
