
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002318
10:01 21/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:18
PRESSURIZING 2:25
STERILIZING 3:27
136.0 C 220kPa 3:27
136.4 C 223kPa 3:57
136.4 C 224kPa 4:27
136.3 C 220kPa 4:57
136.4 C 224kPa 5:27
136.1 C 223kPa 5:57
136.0 C 221kPa 6:27
136.1 C 222kPa 6:57
136.1 C 222kPa 6:57
VENTING 6:57
AIR DRYING 7:47
CYCLE COMPLETE 37:47

Digital Signature #
3F84206CD745E1BA
