
Midmark M11 - v1.0.5
Total Cycles: 2147

Name: RITTER2
Sterilizer ID: V2313228
03/14/2023 11:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.8 F 0.0
2:00 192.4 F 4.2
4:00 213.0 F 9.9
6:00 221.9 F 6.6
8:00 235.1 F 12.2
10:00 237.5 F 10.5
12:00 237.4 F 9.8
14:00 252.4 F 17.1
16:00 264.9 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.5 F 29.5
2:00 271.6 F 29.3
3:00 272.5 F 29.5
4:00 271.5 F 29.0

Min 271.4 F 28.9
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.3 F 0.0
5:00 223.8 F 0.0
10:00 229.4 F 0.0
15:00 235.5 F 0.0
20:00 223.8 F 0.0
25:00 222.8 F 0.0
30:00 219.1 F 0.0
Drying Complete

FILLING: 2:06
HEATING: 17:46
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:57:08
CYCLE COMPLETE

03/14/2023 12:47 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____