
Midmark M11 - v1.0.0
Total Cycles: 1588

Name: RITTER 1
Sterilizer ID: V2178198
02/02/2022 11:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 105.8 F 0.0
2:00 125.9 F 0.4
4:00 136.5 F 2.8
6:00 156.3 F 6.2
8:00 182.4 F 10.2
10:00 208.9 F 4.7
12:00 215.7 F 6.5
14:00 226.8 F 10.8
16:00 231.5 F 7.1
18:00 240.9 F 12.0
20:00 238.9 F 9.8
22:00 251.5 F 16.0
24:00 262.6 F 22.3
26:00 272.8 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 271.7 F 28.7
2:00 272.4 F 28.8
3:00 271.4 F 28.1
4:00 272.2 F 28.9

Min 271.2 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.3 F 0.0
5:00 213.0 F 0.0
10:00 230.9 F 0.0
15:00 221.6 F 0.0
20:00 226.4 F 0.0
25:00 225.0 F 0.0
30:00 221.1 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 26:14
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:04:26
CYCLE COMPLETE

02/02/2022 12:46 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____