
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001024
16:03 06/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CYCLE ABORTED 0:21
STOP BUTTON PRESSED
NOT STERILE
END TIME 0:21

Digital Signature #
11753055BFF68674
