
Midmark M11 - v1.0.0
Total Cycles: 2456

Name: RITTER 1
Sterilizer ID: V2178198
10/19/2022 10:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.3 F 0.0
2:00 184.5 F 4.0
4:00 191.1 F 8.7
6:00 207.5 F 9.9
8:00 222.5 F 8.2
10:00 236.7 F 13.1
12:00 238.6 F 10.9
14:00 236.3 F 9.1
16:00 250.9 F 16.0
18:00 263.1 F 23.0
20:00 271.8 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.5 F 29.2
2:00 271.5 F 28.5
3:00 272.5 F 29.4
4:00 272.0 F 28.8

Min 271.3 F 28.5
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.0 F 0.0
5:00 229.8 F 0.0
10:00 240.5 F 0.0
15:00 240.7 F 0.0
20:00 232.2 F 0.0
25:00 225.6 F 0.0
30:00 226.2 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 20:09
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:36
CYCLE COMPLETE

10/19/2022 11:00 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____