
Midmark M11 - v1.0.5
Total Cycles: 3405

Name: RITTER2
Sterilizer ID: V2313228
02/23/2024 09:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.9 F 0.0
2:00 74.3 F 0.0
4:00 128.1 F 0.0
6:00 99.8 F 0.6
8:00 143.8 F 3.1
10:00 166.6 F 6.4
12:00 185.8 F 9.9
14:00 202.1 F 11.8
16:00 223.6 F 6.9
18:00 233.7 F 10.7
20:00 230.5 F 7.8
22:00 238.6 F 11.3
24:00 235.4 F 8.9
26:00 248.4 F 14.9
28:00 259.0 F 20.9
30:00 268.7 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 272.5 F 29.5
1:00 272.6 F 29.6
2:00 272.5 F 29.7
3:00 272.7 F 29.8
4:00 272.7 F 30.0

Min 271.4 F 29.4
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.1 F 0.0
5:00 231.2 F 0.0
10:00 220.1 F 0.0
15:00 223.7 F 0.0
20:00 216.9 F 0.0
25:00 207.8 F 0.0
30:00 212.5 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 31:42
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:09:51
CYCLE COMPLETE

02/23/2024 10:12 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____