
Midmark M11 - v1.0.0
Total Cycles: 5496

Name:
Sterilizer ID: V2178198
04/10/2025 03:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 132.8 F 0.0
2:00 144.7 F 1.7
4:00 151.8 F 4.9
6:00 177.3 F 8.4
8:00 196.5 F 12.5
10:00 208.7 F 3.6
12:00 220.4 F 7.2
14:00 231.8 F 11.1
16:00 231.5 F 6.7
18:00 241.5 F 11.7
20:00 237.9 F 9.3
22:00 249.9 F 15.1
24:00 260.5 F 21.1
26:00 270.4 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.4 F 29.0
2:00 271.5 F 28.2
3:00 271.9 F 29.0
4:00 272.2 F 28.7

Min 271.1 F 28.2
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.1 F 0.0
5:00 204.4 F 0.0
10:00 192.0 F 0.0
15:00 206.7 F 0.0
20:00 220.5 F 0.0
25:00 223.3 F 0.0
30:00 230.3 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 26:54
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:05:46
CYCLE COMPLETE

04/10/2025 04:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____