
Midmark M11 - v1.0.5
Total Cycles: 920

Name: RITTER2
Sterilizer ID: V2313228
03/29/2022 09:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.0 F 0.0
2:00 175.0 F 0.3
4:00 189.2 F 2.2
6:00 208.6 F 7.3
8:00 230.6 F 13.0
10:00 232.8 F 9.5
12:00 239.0 F 9.3
14:00 243.3 F 12.1
16:00 246.1 F 13.6
18:00 258.8 F 20.5
20:00 270.1 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.6 F 29.4
2:00 271.5 F 28.7
3:00 272.7 F 29.6
4:00 272.0 F 29.0

Min 271.4 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.9 F 0.0
5:00 216.4 F 0.0
10:00 226.0 F 0.0
15:00 229.3 F 0.0
20:00 221.7 F 0.0
25:00 220.5 F 0.0
30:00 220.3 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 20:54
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:57
CYCLE COMPLETE

03/29/2022 10:40 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____