

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002482  
14:24 06/02/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:10  
PRESSURIZING 2:18  
STERILIZING 3:23  
136.0 C 220kPa 3:23  
136.3 C 224kPa 3:53  
136.4 C 225kPa 4:23  
136.3 C 224kPa 4:53  
136.3 C 225kPa 5:23  
136.4 C 226kPa 5:53  
136.2 C 224kPa 6:23  
136.4 C 226kPa 6:53  
136.4 C 226kPa 6:53  
VENTING 6:53  
AIR DRYING 7:47  
CYCLE COMPLETE 23:02  
-----

Digital Signature #  
0CEDA4E992F2F2E5  
-----