

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 003285  
10:50 24/03/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 0:42  
PRESSURIZING 0:51  
STERILIZING 1:53  
136.4 C 225kPa 1:53  
136.2 C 224kPa 2:23  
136.2 C 223kPa 2:53  
136.2 C 223kPa 3:23  
136.2 C 223kPa 3:53  
136.4 C 224kPa 4:23  
136.2 C 223kPa 4:53  
136.4 C 222kPa 5:23  
136.4 C 222kPa 5:23  
VENTING 5:23  
AIR DRYING 6:19  
CYCLE COMPLETE 6:22

-----  
Digital Signature #  
768D8BBED8173BE9  
-----