
Midmark M11 - v1.0.5
Total Cycles: 1813

Name: RITTER2
Sterilizer ID: V2313228
12/06/2022 01:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.6 F 0.0
2:00 192.5 F 4.0
4:00 200.4 F 9.5
6:00 218.5 F 6.6
8:00 233.4 F 10.9
10:00 234.5 F 8.9
12:00 239.6 F 9.8
14:00 249.5 F 15.3
16:00 262.0 F 22.6
18:00 272.6 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.4 F 29.4
2:00 271.7 F 29.2
3:00 272.4 F 29.3
4:00 271.5 F 29.0

Min 271.5 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.7 F 0.0
5:00 221.2 F 0.0
10:00 224.9 F 0.0
15:00 226.6 F 0.0
20:00 226.4 F 0.0
25:00 226.6 F 0.0
30:00 225.6 F 0.0

Drying Complete

FILLING: 2:42
HEATING: 18:15
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:07
CYCLE COMPLETE

12/06/2022 02:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____