
Midmark M11 - v1.0.5
Total Cycles: 3572

Name: RITTER2
Sterilizer ID: V2313228
04/30/2024 01:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 106.1 F 0.0
2:00 109.3 F 1.0
4:00 127.9 F 3.4
6:00 160.6 F 6.6
8:00 187.2 F 10.2
10:00 198.7 F 9.2
12:00 213.4 F 5.4
14:00 225.0 F 9.3
16:00 234.4 F 11.7
18:00 237.5 F 10.7
20:00 235.5 F 8.7
22:00 248.7 F 15.1
24:00 259.8 F 21.3
26:00 270.2 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.8 F 29.7
2:00 272.0 F 29.0
3:00 271.9 F 29.4
4:00 272.8 F 29.6

Min 271.5 F 28.8
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.0 F 0.0
5:00 207.4 F 0.0
10:00 203.0 F 0.0
15:00 199.4 F 0.0
20:00 208.0 F 0.0
25:00 211.6 F 0.0
30:00 213.0 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 26:57
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:05:19
CYCLE COMPLETE

04/30/2024 02:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____