

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 007082  
6:30 22/03/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:21  
PRESSURIZING 3:22  
STERILIZING 4:22  
136.0 C 220kPa 4:22  
136.4 C 220kPa 4:52  
136.3 C 222kPa 5:22  
136.4 C 219kPa 5:52  
136.4 C 219kPa 6:22  
136.3 C 222kPa 6:52  
136.3 C 222kPa 7:22  
136.2 C 218kPa 7:52  
136.2 C 218kPa 7:52  
VENTING 7:52  
AIR DRYING 8:44  
CYCLE COMPLETE 38:44

-----  
Digital Signature #  
817C66E106E873FA  
-----