
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006148
10:17 17/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:37
STERILIZING 3:44
136.4 C 222kPa 3:44
136.4 C 223kPa 4:14
136.4 C 224kPa 4:44
136.4 C 224kPa 5:14
136.4 C 223kPa 5:44
136.4 C 225kPa 6:14
136.4 C 224kPa 6:44
136.4 C 225kPa 7:14
136.4 C 225kPa 7:14
VENTING 7:14
AIR DRYING 8:03
CASSETTE REMOVED
CYCLE COMPLETE 29:57

Digital Signature #
546DEDB626B564CD
