
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 004451
11:05 27/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:37
PRESSURIZING 2:42
STERILIZING 3:43
136.0 C 220kPa 3:43
136.4 C 224kPa 4:13
136.4 C 222kPa 4:43
136.3 C 224kPa 5:13
136.2 C 219kPa 5:43
136.4 C 224kPa 6:13
136.4 C 225kPa 6:43
136.3 C 224kPa 7:13
136.3 C 224kPa 7:13
VENTING 7:13
AIR DRYING 8:05
CYCLE COMPLETE 8:25

Digital Signature #
03BFA887D1BFF9F7
