
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000098
9:55 02/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:26
PRESSURIZING 4:26
STERILIZING 5:50
136.4 C 225kPa 5:50
136.4 C 226kPa 6:21
136.4 C 226kPa 6:51
136.4 C 226kPa 7:21
136.2 C 224kPa 7:51
136.3 C 225kPa 8:21
136.3 C 224kPa 8:51
136.4 C 227kPa 9:21
136.4 C 227kPa 9:21
VENTING 9:21
AIR DRYING 10:21
CYCLE COMPLETE 19:14

Digital Signature #
579C5DF6A6823166
